

URIT HQ 51

HEMATOLOGY CONTROLS

CONTROL

Assay values and ranges provided by URIT Medical
Model skyla SH50



2025/09/21

| Instrument | Parameter | LOT | |
|---|--------------------------------|------------|-------|
| | | 25040667M2 | |
| SH50 | WBC (WOC) × 10 ⁹ /L | 8.79 | ±1.50 |
| | WBC (WIC) × 10 ⁹ /L | 8.13 | ±1.50 |
| | NEU # | 4.04 | ±1.20 |
| | NEU % | 49.62 | ±8.00 |
| | LYM # | 2.68 | ±0.90 |
| | LYM % | 33.00 | ±7.00 |
| Assay values established in QC file. | MON # | 1.17 | ±1.17 |
| | MON % | 14.41 | ±6.50 |
| | EOS # | 0.19 | ±0.19 |
| | EOS % | 2.37 | ±2.37 |
| | BASO # | 0.05 | ±0.05 |
| | BASO % | 0.60 | ±0.60 |
| | RBC × 10 ¹² /L | 3.95 | ±0.25 |
| | HGB g/dL | 11.6 | ±0.7 |
| | HGB g/L | 116 | ±7 |
| | HCT % | 30.6 | ±3.0 |
| | MCV fL | 77.6 | ±7.0 |
| | MCH pg | 29.2 | ±3.6 |
| | MCHC g/dL | 37.8 | ±3.6 |
| | MCHC g/L | 378 | ±36 |
| | RDW-CV % | 15.9 | ±4.0 |
| | PLT × 10 ⁹ /L | 242 | ±50 |
| | MPV fL | 12.6 | ±4.0 |

Note:

1. Take these values for reference only when the quality control material within the validity period. Do not use it as the quality control material expires.
2. Do not take the above values as a reference as the lot number or reagent system being changed.
3. Shake the quality control material up and down for 1-2 minutes as taking it out from refrigerator and place it in room temperature (15°C~30°C) for 10 minutes, and then doing the test.
4. The lowest percent and lowest absolute count of MON, EOS and BASO reference value are 0.

URIT Medical Electronic Co., Ltd.

No. D-07 Information Industry District, High-Tech Zone, Guilin, Guangxi 541004, P.R.China.

Tel: +86(773)2288586 Fax: +86(773)2288560

Web: www.urit.com Email: service@uritest.com

EC-REPRESENTATIVE:

Shanghai International Holding Corp. GmbH (Europe)

Eiffestrasse 80, 20537 Hamburg, Germany

Manufactured for Shenzhen URIT under contract